



CE Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: Laboratory Equipment – Logic (L4) Series Biological Safety Cabinets

Type, Model: 3023zxx-22, 32, 42, 52, 62, 72 Type A2 3' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
 3024zxx-22, 32, 42, 52, 62, 72 Type A2 4' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
 3025zxx-22, 32, 42, 52, 62, 72 Type A2 5' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
 3026zxx-22, 32, 42, 52, 62, 72 Type A2 6' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
 30348xx-22, 32, 42, 52, 62, 72 Type B2 4' Wide, 8" Sash
 30368xx-22, 32, 42, 52, 62, 72 Type B2 6' Wide, 8" Sash
 31232xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 3' Wide 12" Sash
 31242xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 4' Wide 12" Sash
 31252xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 5' Wide 12" Sash
 31262xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 6' Wide 12" Sash
 3y239xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 3' Wide 9" Sash
 3y249xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 4' Wide 9" Sash
 3y259xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 5' Wide 9" Sash
 3y269xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 6' Wide 9" Sash
 3y348xx-22, 32, 42, 52, 62, 72 Type B2 Cell Logic 4' Wide 8" Sash
 3y368xx-22, 32, 42, 52, 62, 72 Type B2 Cell Logic 6' Wide 8" Sash

y = 2 for Temp Zone, 3 for Scope Ready, 4 for Temp Zone + Scope Ready
 x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification. This modification does not impact the compliance to standards listed below.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

are in accordance with the following directives:

2014/30/EU	Electromagnetic Compatibility Directive	2011/65/EU	RoHS Directive
2014/35/EU	Low Voltage Directive	2015/863/EU	RoHS3 Directive
2006/42/EU	Machinery Directive		

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-3-3:2013	EN61000-4-5:2014+A1:2017
EN61326-1:2013	EN61000-4-2:2009	EN61000-4-6:2014
EN55011:2016+A1:2017+A11:2020	EN61000-4-3:2020	EN61000-4-11:2020
EN61000-3-2:2014	EN61000-4-4:2012	EN63000:2018

when installed and operated in accordance with the manufacturer's installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

11 DEC 2023
Date of issue



UKCA Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:

Equipment: Laboratory Equipment – Logic (L4) Series Biological Safety Cabinets

Type, Model:

3023zxx-22, 32, 42, 52, 62, 72	Type A2 3' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3024zxx-22, 32, 42, 52, 62, 72	Type A2 4' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3025zxx-22, 32, 42, 52, 62, 72	Type A2 5' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3026zxx-22, 32, 42, 52, 62, 72	Type A2 6' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
30348xx-22, 32, 42, 52, 62, 72	Type B2 4' Wide, 8" Sash
30368xx-22, 32, 42, 52, 62, 72	Type B2 6' Wide, 8" Sash
31232xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 3' Wide 12" Sash
31242xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 4' Wide 12" Sash
31252xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 5' Wide 12" Sash
31262xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 6' Wide 12" Sash
3y239xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 3' Wide 9" Sash
3y249xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 4' Wide 9" Sash
3y259xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 5' Wide 9" Sash
3y269xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 6' Wide 9" Sash
3y348xx-22, 32, 42, 52, 62, 72	Type B2 Cell Logic 4' Wide 8" Sash
3y368xx-22, 32, 42, 52, 62, 72	Type B2 Cell Logic 6' Wide 8" Sash

y = 2 for Temp Zone, 3 for Scope Ready, 4 for Temp Zone + Scope Ready
x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification.
This modification does not impact the compliance to standards listed below.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

S.I. 2016/1091	Electromagnetic Compatibility Directive	S.I. 2008/1597	Machinery Directive
S.I. 2016/1101	Low Voltage Directive	S.I. 2012/3032	RoHS Directive

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-3-3:2013	EN61000-4-5:2014+A1:2017
EN61326-1:2013	EN61000-4-2:2009	EN61000-4-6:2014
EN55011:2016+A1:2017+A11:2020	EN61000-4-3:2020	EN61000-4-11:2020
EN61000-3-2:2014	EN61000-4-4:2012	EN63000:2018

when installed and operated in accordance with the manufacturer's installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

11 DEC 2023
Date of issue