

Product Data Sheet

23 CF Premium Laboratory Glass Door Refrigerator

ABT-HC-PL-23

Product Descriptio

ABS Premium Lab Refrigerators deliver superior cooling to laboratory environments. Engineered with variable speed compressors (VSCs), these units feature ultra-quiet operation and significant energy savings. VSCs optimize energy consumption by self-adjusting to cooling demands. These compressors also deliver enhanced system performance and provide a longer lifespan than other compressor variations.

Enjoy the uniformity and speed of microprocessor temperature control and a full array of alarms to safeguard your products. Unit controllers also come with a battery backup to keep your data safe. Upgrade your laboratory environment with these premium refrigerators and welcome energy savings, noise reduction, smooth operation, and improved system performance to your work environment.

Images




Certifications




neral Description and Application	
Storage capacity (cu. ft)	23
Door	Single Swing Glass Right Hinged Door(s)
Shelves	Five adjustable shelves with guard rail on back
Drawers	Optional pull-out drawers available
Mounting and Installation	4 preinstalled swivel casters, front casters locking
Interior lighting	Shielded, switched LED lighting, full coverage, balanced spectrum
Airflow Management	Forced Air technology, patent pending
External probe access	Rear wall port (3/4") dia.
Insulation	Cabinet is foamed-in-place with EPA compliant high density urethane foam
Exterior materials	White powder coated steel
Access control	Key Lock
General warranty	Two (2) years parts and labor warranty
Compressor warranty	Seven (7) years compressor warranty
Product Weight (lbs)	350
Shipping Weight (lbs)	409
Rated Amperage	3 Amps
Power Plug/Power Cord	NEMA 5-15 plug
Facility Electrical Requirement	110-120V AC: 15 A (minimum)
Agency Listing and Certification	ETL, C-ETL listed and certified to UL471 standard, hydrocarbon refrigerant safety


rformance	
Uniformity ¹ (Cabinet air)	±0.6
Stability ² (Cabinet air)	±0.6
Maximum temperature variation	±0.9
(Cabinet air)	
Stability ² (Simulator ballast)	Non-applicable
Stability ² (Simulator bag)	Non-applicable
Temperature Rise after Short Door Openings	Non-applicable
Recovery after Short Door Openings	Non-applicable
Energy Consumption (KWh/day)	1.15
Average Heat Rejection (BTU/hr)	320
Noise Pressure Level (dBA)	35 or less installed
Pull down time to nominal operating temp	55 min

3.8 3.2 4.3 4.3 3.9 4.6 4.3 4.6 3.9 4.4 4.0 4.7 4.3 4.0 4.6 3.8 3.3 4.4 4.2 4.7 3.8 4.6 3.8 4.9 4.4 3.8 4.8 4.4 4.7 10 4.0 11 4.0 3.3 4.5 12 3.9 3.4 4.3 4.2 3.8 13 4.5 14 4.5 4.0 4.9 15 3.8 3.2 4.3 Bag N/A N/A



Temperature Charts






Performance data acquired at 22°C ambient, 4°C nominal set point in an empty cabinet with drawers using validation ballast probes, during stabilized steady state operation and a DAQ sampling rate of one measurement every 10 seconds


- 1 Uniformity is defined as the maximum variance in temperature across all probes at any point in time over the testing period
- 2 Stability is defined as the maximum variance in temperature experienced by any single probe over the testing period
- 3 Data per Energy Star test results or equivalent testing and calculation. Heat rejection based on daily overages, not continuous operation. Performance exceeds Energy Star requirements.

Refrigeration System		
Compressor	Hermetic, variable speed (VSC). Rated speed range: 1300-4000 rpm	
Refrigerant	EPA SNAP compliant, R600a	
Condenser	Anti-fouling tube and grid design, ultra-quiet multi-speed fan	
Evaporator	Fin and tube design, high efficiency fan	
Defrost	Cycle optimized, zero energy	

o energy I Derivative (PID) microprocessor with LCD display ttery, controller, all alarms active, temperature monitoring DAQ and event logging active on battery backup
ittery, controller, all alarms active, temperature monitoring DAQ and event logging active on battery backup
insfer and software updates
ote alarm contacts
00 entries) with Visual and audible indicators: Power failure, Temperature sensor failure, Battery voltage monitor and replacement, High / Low temperature, Door ajar.
d thermal media
equired at 22°C ambient, 4°C nominal set point in an empty cabinet with drawers using validation ballast probes, during stabilized steady state operation and a DAQ measurement every 10 seconds ned as the maximum variance in temperature across all probes at any point in time over the testing period d as the maximum variance in temperature experienced by any single probe over the testing period tar test results or equivalent testing and calculation. Heat rejection based on daily averages, not continuous operation. Performance exceeds Energy Star requirements.
on on

Dimensions					
	Width (in.)	Depth (in.)	Height (in.)	Door Swing (in.)	Total open Depth (in.)
Exterior	26 7/8"	35 3/4"	81 1/2"	25 1/4"	59 1/8"
Interior	22"	25"	52 1/4"		





Contact		
Customer Service	800-648-4041 Option 3	<u>customerservice@horizonscientific.com</u>
Technical Service	800-648-4041 Option 5, Parts Option 4	technicalservice@horizonscientific.com
Rev_06162023		