

# Cool Soak™

## Stain & Rust Remover



Cool Soak™ Stain & Rust Remover is specially formulated to remove rust, water spots and detergent residue from stainless steel and tungsten carbide surgical instruments; automatic washing equipment and autoclaves. Cook Soak may be used manually, in ultrasonic washers and for descaling automatic washing equipment and autoclaves.

- **Restores instruments' luster and shine.**
- **Keeps box locks moving freely.**
- **Safe for stainless steel and tungsten carbide instruments.**

Order No.	Size	Physical Properties
T5	1 Gallon (3.8 Liters)	<b>Form:</b> Liquid <b>Color:</b> Violet <b>Odor:</b> Mild <b>pH Concentration:</b> <1

Uses	Temperature
<input type="checkbox"/> Presoak <input checked="" type="checkbox"/> Automatic Washer <input checked="" type="checkbox"/> Manual <input type="checkbox"/> Automatic Endoscope Reprocessor <input checked="" type="checkbox"/> Ultrasonic <input checked="" type="checkbox"/> Cart Washer	<b>Usage:</b> Room Temperature <b>Storage:</b> Cool, dry area

Product Features	Consumer Benefits
<b>Removes Stains and Rust:</b>	Restores instruments' luster and shine.
<b>Dissolves Residues:</b>	Keeps box locks moving freely.
<b>Instrument Safe:</b>	If used properly, does not harm stainless steel or tungsten carbide instruments.
<b>Bacteriostatic:</b>	If stored properly, not favorable to bacterial growth.

Procedures	Directions for Use
<b>Directions:</b>	Wear appropriate personal protective equipment based on safe work practices.
<b>Manual Descaling:</b>	Dilute 1 part Cool Soak with 7 parts warm water. Soak thoroughly cleaned and opened instruments in prepared solution for 5-15 minutes. Scrub instruments under water with the Cool Soak solution and rinse thoroughly with water. Thoroughly clean instruments again once Cool Soak has been used and then lubricate and sterilize as normal. Change solution daily or more often as needed.
<b>Ultrasonic Descaling:</b>	Dilute 1 part Cool Soak with 7 parts water.
<b>Descaling Automatic Washing Equipment:</b>	Dilute 1 fl. oz. (30 mL) to 3 fl. oz. (88 mL) Cool Soak per gallon (3.8 L) of water and follow manufacturer's guidelines for descaling.
<b>Descaling Autoclaves:</b>	Dilute 1 part Cool Soak with 1 part water. Apply solution to cool chamber walls and soak for 10-15 minutes. Wipe off (scrubbing may be necessary for heavier deposits) and rinse thoroughly with warm water.