

KIMTECH™ Element™ Nitrile Exam Glove

Blue nitrile, powder-free, non-sterile ambidextrous exam glove with beaded cuff and textured fingertips. Provides wearer with protection against blood and bodily fluids, as well as other liquids and particulates encountered during a medical examination. These gloves comply with FDA 21 CFR 177.2600 for rubber in contact with food. Layer-stacked and easily dispensed.

Product Information

Code	Size	Palm Width* (mm)	Length* (mm)	Qty per dispenser	Qty per case
62870	XS	75	237	250	2500
62871	S	85	237		
62872	M	95	237		
62873	L	105	237		
62874	XL	115	237	230	2300

*As measured by ASTM D6319, AQL 4.0

Physical Property	Unit	Target	Test Method
Residual powder	mg/glove	≤2	ASTM D6124
Watertightness	—	Pass	ASTM D5151

Thickness	Target (mm)	Target (mil)	Test Method
Finger	0.08	3.15	ASTM D6319
Palm	0.05	1.97	
Cuff	0.04	1.58	

Property	Unit	Target	Test Method (4.0 AQL)
Tensile strength before aging	MPa	≥14	ASTM D6319
Tensile strength after aging	MPa	≥14	
Ultimate elongation before aging	%	≥500	
Ultimate elongation after aging	%	≥400	



Workmanship Requirement: Gloves shall be made in accordance with reasonable industry practice with respect to defects, dirt and contamination. Gloves should be packaged as specified having described size and color in the appropriate quantity as listed. Shipping Case and inner packaging shall be labeled with product name, size, lot/batch and catalog number.

For Additional information: Contact your Kimberly-Clark representative; our North America Customer Service Team (888)-346-GOKC (4652), KCPIInfo@kcc.com; or our Asia-Pacific (AP) Customer Service Team kimtech.asia@kcc.com.

KCC Confidential: The information contained in this document is intended for the use of the addressee(s) only and may contain privileged, confidential, or proprietary information that may not be distributed publicly without Kimberly-Clark's written consent.

Certificate of Conformance available online at www.kimtech.com/certificates.

Revised July 28, 2021

Kimberly-Clark warrants that its products (1) comply with K-C's standard specifications as of the delivery date to K-C's authorized distributors/direct purchasers (2) comply with all K-C labeling representations; and (3) are manufactured in compliance with all applicable federal, state, and local laws in effect at the time and place of manufacture of the products. **THIS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** K-C is not liable for any kind of special, incidental, or consequential damages. K-C's liability for breach of contract, tort or other cause of action shall not exceed the product purchase price. Purchasers and users are deemed to have accepted the above warranty and limitation of liability, and cannot change the terms by verbal agreement or by any writing not signed by K-C. To the extent required by applicable law, K-C does not limit its liability for death/injury resulting from K-C's negligence.

™/® Trademarks of Kimberly-Clark Worldwide, Inc., or their respective holders and may be used under license. Marques de commerce de Kimberly-Clark Worldwide, Inc. ou de leurs détenteurs respectifs et peuvent être utilisées sous licence © KCWW.

